










Symbol	Standard Reference Number	Standard Title	Symbol Title	Description of Symbol
	ISO 15223-1, Clause 5.1.1	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied.	Manufacturer	Indicates the medical device manufacturer.
	ISO 15223-1, Clause 5.1.5	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied	Batch code	Indicates the manufacturer's batch code so that the batch or lot can be identified.
	ISO 15223-1, Clause 5.1.4	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied.	Use by	Indicates the date after which the medical device is not to be used.
	ISO 15223-1, Clause 5.4.3	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied.	Consult instructions for use	Indicates the need for the user to consult the instructions for use.
	ISO 15223-1, Clause 5.4.4	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied.	Attention: Read all warnings and precautions in instructions for use	Indicates the need for the user to consult the instructions for use for important cautionary information such as warnings and precautions that cannot, for a variety of reasons, be presented on the medical device itself.

Confidential

Do not reproduce without written consent from Pelvalon, Inc.

	ISO 15223-1, Clause 5.4.2	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied.	Do not reuse	Indicates a medical device that is intended for one use, or for use on a single patient during a single procedure.
	ISO 15223-1, Clause 5.2.7	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied.	Non sterile	Indicates a medical device that has not been subjected to a sterilization process.
	ISO 15223-1, Clause 5.1.6	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied.	Catalogue or model number	Indicates the manufacturer's catalogue number so that the medical device can be identified.
	21 CFR 801.15(c)(1)(i)F	Labeling-Medical devices; prominence of required label statements.	Prescription only	Requires prescription in the United States.
	21 CFR 801.109	Labeling-Prescription devices.		

1. REVISION HISTORY:

Rev	Effective Date	Change Description	Originator	ECO #
A	08/28/17	Initial Release	G.Tejada	ECO0340

Confidential

Do not reproduce without written consent from Pelvalon, Inc.